




Stock Code: 688068.SH

Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold) Summary Data



Beijing Hotgen Biotech Co., Ltd.


目录

(Contents)

1、产品彩页 (Product Brochure)	1-2
2、公司介绍 (Company Profile)	3
3、CE 申明 (Declaration of Conformity)	4
4、说明书 (Instructions for Use)	5-6
5、产品照片 (Product Photos)	7
6、包装信息 (Packing Information)	8
7、临床验证报告 (Clinical Trial Summary Report)	9-12
8、检测灵敏度 (The Sensitivity of Test)	13
9、航空鉴定书 (Certification for Safe Transport of Chemical Goods)	14
10、ISO13485 认证 (ISO13485 Certificate)	15
11、企业资质 (Enterprise Qualification)	16-18



Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)



Product Features

- High Accuracy, Specificity and Sensitivity ●
- No need instrument, get results in 15 minutes ●
- Room Temperature Storage ●
- Sample : Nasopharyngeal swab , Throat Swab ●
- Detect the presence of viral proteins ●
- Identify acute or early infection ●
- The sensitivity is 2.5×10^2 pfu/mL ●


Clinical Performance

(Disease Course 5-7 Days)


Sensitivity: 95.65%; Specificity: 99.02%; Accuracy: 97.89%.

Novel Coronavirus 2019-nCoV Antigen Test(Colloidal Gold)

Sample Collection




Nasopharyngeal swab: The sampler holds the swab to enter the nostril, and when the tip of the swab reaches the back wall of the nasopharyngeal cavity, gently rotate it for a circle, and then slowly take out the swab.



Throat Swab: The swab crosses the base of the tongue, and wipes the tonsils back and forth with slight force on both sides of the person being collected for at least 3 times, and then wipes up and down the posterior pharyngeal wall at least 3 times.

Test Procedure



1 The swab after sampling is soaked below the liquid level of the sampling tube, rotated and pressing 3 times, the swab soaking time is not less than 15s, the swab head is pressed, then taken out the swab and tighten the sampling tube. The liquid in the tube is the sample after treatment.






2 Add 4-5 drops of the treated sample into the sample well of the test cassette.



3 Observe results after 15 minutes, result got after 30 minutes is invalid.

Interpretation of result



Clinical Performance

A total of 617 nasal swab samples were tested in this test, and the results of throat swabs samples were analyzed statistically. The collecting time of patient samples is not exceeding 7 days after clinical manifestations with a novel coronavirus infection in clinical institutions.

Assessment system Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)	Reference system (clinical diagnostic results)		
	Positive(+)	Negative(-)	Total
Positive(+)	198	4	202
Negative(-)	9	406	415
Total	207	410	617

Sensitivity: 95.65%; Specificity: 99.02%; Accuracy: 97.89%.

Product information

Product name	Test samples	Specifications	Storage conditions
Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)	Nasopharyngeal swab , Throat Swab	1T/kit, 5T/kit, 20T/kit, 40T/kit	4-30 C

Company Profile

Beijing Hotgen Biotechn Co., Ltd. (abbreviated as Hotgen Biotech, stock code: 688068) was established in June 2005, which is a high-tech enterprise focusing on the research& development, manufacture and sales of medical and public safety inspection products of in vitro diagnostics (IVD) in the field of biomedicine, as well as landed on the China Sci-Tech innovation board (STAR Market) in September 2019.

After several years of Research& development, Hotgen Biotech has developed an in vitro diagnostic reagent bioactive raw material development platform, a sugar chain abnormal protein detection (sugar capture) R&D technology platform, a Magnetic particles chemiluminescence R&D technology platform, a Up-converting Phosphor R&D technology platform, and a colloidal gold immune layer, The eight major technology platforms, such as the precipitation R&D platform, enzyme-linked immunoassay R&D technology platform, molecular diagnostics R&D platform, and instrument R&D technology platform, form a closed-loop system for in vitro diagnostic R&D and production. Hotgen Biotech has established a complete full level immunodiagnostic technology platform, from high-precision Up-converting Phosphor POCT (UPT series) to small, medium and large single- cartridge chemiluminescence platforms (MQ60 series), and then to large-scale full-automatic chemiluminescence Platform (C2000), which realizes the application of the immune diagnostic platform in the field of full diagnostic scenarios. Supporting products are widely used in the clinical and public safety fields. Specific users include hospitals at all levels, township health centers, third-party testing centers, and medical institutions, as well as medical and health institutions, as well as disease control centers, public security, fire protection, military, ports, food and medicine. Supervision, food and feed enterprises and other public safety fields.

Hotgen Biotech has won the second prize of the National Technology Invention Award, the Gold Medal of Independent Innovation, and the second prize of the Chinese Medical Science and Technology Award; In 2018, Hotgen Biotech was awarded the second prize of the "Technical Invention Category of China Rare Earth Science and Technology Award" by the China Rare Earth Society; Top 100 Private Scientific and Technological Innovations "and" Top 100 Medical Enterprises of the Future "; and" Postdoctoral Scientific Research Workstation "; major science and technology projects in the 12th and 13th five years, 863 plan, science and technology projects of the Beijing Science and Technology Commission, and Zhongguancun High Precision The project's major cutting-edge original technological achievements transformation and industrialization projects.

In the face of the COVID-19 epidemic situation, Beijing Hotgen Biotech Co.,Ltd has organized R&D developed a variety of Covid-19 detection reagents, including Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold), Novel Coronavirus 2019-nCoV Antigen Test (Up-converting Phosphor Immunochemical Technology), Coronavirus disease(COVID-19) Antibody Test (Colloidal Gold), Coronavirus disease(COVID-19) IgM/IgG Antibody Rapid Test (Colloidal Gold), Novel Coronavirus 2019-nCoV Antibody Test (Up-converting Phosphor Immunochemical Technology), Novel Coronavirus (SARS-CoV-2) Neutralizing Antibodies Test (Colloidal Gold), Coronavirus disease(COVID-19) Nucleic Acid Test Kit (PCR-Fluorescent Probe Method), Disposable virus sampling tube, Nucleic acid Automatic Purification System, Nucleic acid extraction reagent, Biological Sample Releaser kit, etc.It is imperative to fight the epidemic Helping the global fight against epidemics!

Since its establishment, the company has continuously grown its business and has now achieved group development. At present, Hotgen (Langfang), Hotgen (Jilin), Weikekang Technology, Shunjing Biological and many other subsidiaries have been established. Hotgen Biotech marketing and service network has covered all regions of the country. Each province is equipped with professional technical service engineers, academic engineers, etc. who are responsible for pre-sales and after-sales work to meet customer needs. The company takes "developing biotechnology and benefiting human health" as its mission, "quality determines the company's life and death, customers determine the company's success or failure, talents determine the company's rise and fall, innovation determines the company's future" as its core values, and "tests because of me advanced" as its philosophy , High ambitions, technological entrepreneurship, and industrial prosperity!

Declaration of Conformity

Manufacturer:

Name: Beijing Hotgen Biotech Co.,Ltd

Address: 9th building, No.9 Tianfu Street, Biomedical Base,Daxing District, Beijing,
102600, P.R.China

European Representative:

MedNet GmbH

Borkstrasse 10,48163 Muenster,Germany

Product Name:

Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)

Novel Coronavirus 2019-nCoV Antigen Test (Up-converting Phosphor Immunochromatographic
Technology)

Classification : **Others of ANNEX II of IVDD**

Conformity Assessment Route: **Annex III**

We herewith declare that the above mentioned products meet the transposition into national law, the provisions of the following EC Council Directives and Standards. All supporting documentations are retained under the premises of the manufacturer.

General applicable directives:

In Vitro Diagnostic Medical Devices DIRECTIVE 98/79/EC

Harmonized standards:

EN ISO 13485:2016, EN ISO 15223-1:2016, EN ISO 14971:2012,EN 13975:2003,
EN ISO 18113-1:2011, EN ISO 18113-2:2011,EN 13612:2002,EN ISO 17511:2003,
EN ISO 23640:2015, EN 13641:2002,EN 13975:2003, EN 62366:2008



Signature: *Lin Changqing*

Name: Lin Changqing

Title: General manager

Place: Beijing,China.

Date of Issue: Aug 27, 2020

Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)

Instructions for Use

PRODUCT NAME

Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)

SPECIFICATIONS

1T/bag, 5T/kit, 20T/kit, 40T/kit

INTENDED USE

This kit is used for in vitro qualitative determination of novel coronavirus antigen in human nasal swabs or throat swabs. It is used as rapid investigation for suspected cases of novel coronavirus, can also be used as a reconfirmation method for nucleic acid detection in discharged cases.

A positive test result indicates that the samples contained novel coronavirus antigen. A negative test result does not rule out the possibility of infection.

This product is only used for clinical and emergency reserve during the pneumonia outbreak of new coronavirus infection, and can not be used as a routine in vitro diagnostic reagent for clinical application. The test results of this kit are for clinical reference only. It is recommended to conduct a comprehensive analysis of the condition based on the patient's clinical manifestations and other laboratory tests.

PRINCIPLE OF THE ASSAY

This kit is based on the Colloidal gold immunochromatographic technology, and uses double antibody sandwich method to detect the novel coronavirus antigen in human throat swabs or nasal swabs. The detection line (T line) of the novel coronavirus antigen test cassette was coated with novel coronavirus antibody, and the quality control line (C line) was coated with sheep anti-mouse. During the test, the sample is dropped into the test cassette and the liquid is chromatographed upward under the capillary effect. The novel coronavirus antigen in the sample first binds to the Colloidal gold-labelled novel coronavirus antibody to form a solid phase novel coronavirus antibody-novel coronavirus antigen-labelled novel coronavirus antibody-Colloidal gold complex at the T line position, and form a solid phase sheep anti-mouse-labelled novel coronavirus antibody- Colloidal gold complex was formed at the C line position. After the test is completed, observe the Colloidal gold color reaction of T line and C line to determine results of novel coronavirus antigen in nasal swabs or throat swabs.

COMPONENTS

1. Novel Coronavirus Antigen Test Cassette 2. Sample extraction buffer 3. Disposable virus sampling swab

STORAGE AND SHELF LIFE

- The kit should be stored at 4~30°C, the shelf life is set for 18 months.
- After the foil bag is opened, it should be used within 30 minutes (temperature 10~30°C, humidity ≤70%), and it should be used immediately after opening at 30°C.
- The sample extraction buffer should be used within 18 months after opening (temperature 10~30°C, humidity ≤70%).
- Date of manufacture and expiration date see label.

SPECIMEN REQUIREMENTS

1. Sample collection:

Nasal swab: The sampling staff hold a swab and stick into the nostril and goes back slowly along the bottom of the lower nasal canal, when the top of the swab reaches the posterior wall of the nasopharyngeal cavity, rotate gently for a

cycle (if reflex cough, stay for a moment), and then slowly remove the swab.

Throat swab: Let the patient's head tilt slightly, mouth open, and make "ah" sounds, exposing the pharyngeal tonsils on both sides. Hold the swab and wipe the pharyngeal tonsils on both sides of the patient with a little force back and forth at least 3 times. Then wipe up and down the Posterior pharyngeal wall at least 3 times.

2. Sample treatment

The swab after sampling is soaked below the liquid level of the sampling tube, rotated and pressing 3 times, the swab soaking time is not less than 15s, the swab head is pressed, then taken out the swab and tighten the sampling tube. The liquid in the tube is the sample after treatment.

3. Sample preservation

The sample of treated should be tested within 1h. Specimens that can not be detected within 24 hours should be kept at -70°C or below. Repeated freezing and thawing should be avoided during specimen transportation. Specimen collection should be sent to the laboratory as soon as possible. If it is necessary to transport the specimen for a long distance, it is recommended to preserve the specimen by refrigeration such as dry ice.

TEST PROCEDURE

- Place the test cassette, sample extraction buffer at room temperature for 15~30 minutes, and equilibrate to room temperature (10~30°C).
- Open the aluminum foil pouch of the test cassette, place the test cassette on a flat surface.
- Write sample ID on the plastic case of the test cassette.
- Add 4~5 drops of the treated sample into the sample well of the test cassette. Incubate at 10~30°C for 15 minutes.
- Observe the results after Incubate at 10~30°C for 15 minutes. Result got after 30 minutes is invalid.


This kit doesn't have quality control products. It is recommended that the users establish a quality control method suitable for its laboratory.

INTERPRETATION OF RESULT

Positive: Two color bands appear in the observation window, that is, a red or magenta line appears at the position of the quality control line (C line) and the detection line (T line) (as shown in result 1), which indicates the test result of novel coronavirus antibody in the sample was positive.

Negative: A red or magenta line appears at the position of the quality control line (C line) in the observation window, and no line appears at the position of the test line (T line) (as shown in the result 2), indicating the test results of the novel coronavirus antibodies in the sample were negative or the concentration was below the limit of detection of the kit.

Invalid: No line appears in the position of the quality control line (line C) in the observation window (as shown in result 3), which indicates that the test is invalid, should collect sample again and retest.

**LIMITATIONS**

1. This kit is a qualitative test and cannot quantify the concentration of the novel coronavirus antigen.
2. The test result of this kit is not the only confirmation indicator of clinical indications. If the test result is not in consistent with clinical evidence, it is recommended to conduct supplementary tests to verify the result.
3. Sample test results are related to the quality of sample collection, processing, transportation and storage. Any errors may cause inaccurate test results. If cross-contamination is not controlled during sample processing, false positive results may occur.

PERFORMANCE CHARACTERISTICS

1. When testing with enterprise references, meet the following standards:
 - 1.1 Negative references compliance rate: Use the enterprise negative references for testing, and the negative references should be detected at least 20/20 (-/-).
 - 1.2 Positive references compliance rate: Use the enterprise positive references for testing, and the positive references should be detected at least 5/5 (+ / +).
 - 1.3 Sensitivity references: When using enterprise sensitivity references for detection, at least 1/3 (+ / +) should be detected.
 - 1.4 Repeatability: Use enterprise precision references for testing, and the test results of repeatable references should be consistent.

PRECAUTIONS

1. This kit is for in vitro diagnostic use only. Please read this instruction carefully before experiment.
2. Please use the swab and sample extraction buffer provided by this kit, Do not replace the sample extract in this kit with components in other kits.
3. Operation should be strictly performed according to the instruction, and different batches should not be mixed use.
4. The user should test the specimen as soon as possible, and the clinical performance evaluation of frozen sample may be different from that of fresh sample.
5. Positive and negative predictive values are highly dependent on prevalence rates. Positive test results are more likely to represent false positive results during periods of little/no SARS-CoV-2 activity when disease prevalence is low. False negative test results are more likely when prevalence of disease caused by SARS-CoV-2 is high.
6. Sensitivity of the test after the first five days of the onset of symptoms has been demonstrated to decrease as compared to a RT-PCR SARS-CoV-2 assay.
7. The test cassette must be used within 30 minutes after opening (temperature 10~30°C, humidity ≤70%), it should be used immediately after opening at 30°C, and the unused test cassette must be sealed and dryly stored.

8. Waste or excess samples produced during testing should be inactivated according to infectious agents.

EXPLANATION FOR IDENTIFICATION

	Use by date	LOT	Batch		Consult Instruction for use
	Content Sufficient For <n> Tests		Temperature limitation	REF	Catalog Number
	Manufacturing date		Caution		Do not reuse
CE	CE Marking – IVDD 98/79/EC	EC REP	Authorized representative in the European Community		Manufacturer
IVD	For In Vitro Diagnostic Use	/	/	/	/

REFERENCES

- [1] Li M, Jin R, Peng Y, et al. Generation of antibodies against COVID-19 virus for development of diagnostic tools [J]. medRxiv. <https://doi.org/10.1101/2020.02.20.20025999>.
- [2] Diagnosis and Treatment Protocol for Novel Coronavirus Pneumonia (Trial Version VII), National Health Office Medical Letter [2020] No.184, 2020.3.3.
- [3] Laboratory Biosafety Guidance Related to Coronavirus Disease (COVID-19) (Edition II), National Health Office Science and Education Letter [2020] No. 70, 2020.1.23.
- [4] Technical Guidelines for COVID-19 Laboratory Testing (Edition IV), National Health Office's Disease Control Letter [2020] No.109, 2020.02.07.



Beijing Hotgen Biotech Co., Ltd.
9th Building, No. 9 Tianfu Street, Biomedical Base, Daxing District, Beijing, 102600, P.R. China.






MedNet GmbH
Borkstrasse 10, 48163 Muenster, Germany



Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)

Product Photos



抗原胶体金检测试剂包装信息

Novel Coronavirus 2019-nCoV Antigen Test(Colloidal Gold) (Colloidal Gold)

Packing Information

产品名称 Product name	规格/盒 Specifications	单位 Unit	单位包装毛重 Gross weight per unit package
Novel Coronavirus 2019-nCoV Antigen Test(Colloidal Gold)	1T	盒/kit	0.039kg/盒 0.039kg / kit
	20T	盒/kit	0.884 kg/盒 0.884 kg / kit

抗原胶体金试剂盒出口包装箱

Novel Coronavirus 2019-nCoV Antigen Test(Colloidal Gold) Export Packing Cartons

包装箱/ 盒 Packing Carton/ box	长 length cm	宽 Width cm	高 height cm	规格 Specifi cation s	每箱装盒 数量 Kit quantity per carton	单盒试剂 净重 Net weight of single kit	整箱净重 Net weight of the whole carton	抛重 Throwing weight
纸箱 carton	70.5	40	39	1T	320 盒 320 kits	0.039 公斤 0.039kg	12.48 公斤 12.48 kg	18.5-19 公斤/kg
纸箱 carton	70.5	40	39	20T	16 盒 16 kits	0.884 公斤 0.884 kg	14.14 公斤 14.14 kg	18.5 - 19 公斤/kg

Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)

Clinical Trial Summary Report

Research product name: Novel Coronavirus 2019-nCoV

Antigen Test (Colloidal Gold)

Test start time: May 6th,2020

Test completion time: Aug.13th,2020

Model specifications: 40T/kit

Medical institutions undertaking clinical trials:

Fifth Medical Center of General Hospital of Chinese People's
Liberation Army

The Sixth People's Hospital of Shenyang

Institute of Microbiology and Epidemiology, Academy of Military
Medical Sciences

Peking Union Medical College Hospital, Chinese Academy of
Medical Sciences

PLA Third Medical Center

Applicant: Beijing Hotgen Biotech Co., Ltd.


Reporting time: Aug.17th,2020

Beijing Hotgen Biotech Co., Ltd.

Summary of Research Report

Clinical trial sponsor	Beijing Hotgen Biotech Co., Ltd.
Clinical trial name	Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)
Clinical trial facility	Fifth Medical Center of General Hospital of Chinese People's Liberation Army, The Sixth People's Hospital of Shenyang, Institute of Microbiology and Epidemiology Academy of Military Medical Sciences, Peking Union Medical College Hospital Chinese Academy of Medical Sciences, PLA Third Medical Center
Purpose of clinical trials	Examine the clinical performance of the Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold) for the detection of novel coronavirus 2019-nCoV antigen in human nasal swabs or throat swabs.
Clinical trial methods	In this clinical trial, the diagnostic criteria for the diagnosis of Coronavirus disease (COVID-19) infection and the results of the disease process (real-time fluorescent RT-PCR detection of novel coronavirus 2019-nCoV nucleic acid results, virus gene sequencing comparison) were selected as comparative methods for comparative research. Test results on clinical case samples. Statistics and calculation of the detection coincidence rate of the two. The differential samples should be fully analyzed in combination with the patient's epidemiological background, clinical symptoms, and disease outcomes to assess the Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold) produced by Beijing Hotgen Biotech Co., Ltd. is used to qualitatively test the clinical performance of the novel coronavirus 2019-nCoV antigen in human nasal swabs or throat swabs.
Test kit name, specifications, batch number	Name: Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold) Specification: 40 Tests/Kit; Lot number: W2020040300
Sample size	The total number of nasal swabs samples was 207 cases of NDV positive samples and 410 cases of NDV negative samples; The total number of throat swabs samples was 201 cases of NDV positive samples and 402 cases of NDV negative samples; Negative nasal swabs samples included 10 cases of HBsAg positive, 7 cases of HCV positive, 2 cases of HIV positive, 6 cases of abnormal liver function, 7 cases of abnormal renal function, 3

	<p>cases of abnormal blood glucose, influenza A, influenza B, and mycoplasma pneumoniae, Fever, upper respiratory tract infection, viral hepatitis, cirrhosis, brucellosis, etc. Negative throat swabs samples included 9 cases of HBsAg positive, 6 cases of HCV positive, 2 cases of HIV positive, 5 cases of abnormal liver function, 6 cases of abnormal renal function, 3 cases of abnormal blood glucose, influenza A, influenza B, and mycoplasma pneumoniae, Fever, upper respiratory tract infection, viral hepatitis, cirrhosis, brucellosis, etc.</p>																									
<p>Judgment method</p>	<p>Visual observation</p>																									
<p>Evaluation method</p>	<p>(1) The total coincidence rate of the diagnosis results of the assessment system and the reference system is greater than 80%. (2) The Kappa value of the consistency between the diagnostic results of the assessment system and the reference system is greater than 0.75.</p>																									
<p>Results and conclusions</p>	<p>1、 The sensitivity , spesitivity , and accuracy of the diagnostic results of the assessment system and the reference system are: Nasal swabs samples, 95.65%, 99.02%, and 97.89% Throat swabs samples,96.02%,98.51%,97.69% (1) human nasal swabs</p> <table border="1" data-bbox="443 1234 1463 1529"> <thead> <tr> <th rowspan="2">Assessment system Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)</th> <th colspan="3">Reference system (clinical diagnostic results)</th> </tr> <tr> <th>Positive (+)</th> <th>Negative (—)</th> <th>Total</th> </tr> </thead> <tbody> <tr> <td>Positive (+)</td> <td>198</td> <td>4</td> <td>202</td> </tr> <tr> <td>Negative (—)</td> <td>9</td> <td>406</td> <td>415</td> </tr> <tr> <td>Total</td> <td>207</td> <td>410</td> <td>617</td> </tr> </tbody> </table> <p>Sensitivity: 95.65% Spesitivity: 99.02% Accuracy: 97.89% Confidence interval analysis with a total compliance rate of 95%:</p> <table border="1" data-bbox="432 1738 1463 1827"> <thead> <tr> <th>Total compliance</th> <th colspan="2">95% confidence interval</th> </tr> </thead> <tbody> <tr> <td>97.89%</td> <td>96.42%</td> <td>98.87%</td> </tr> </tbody> </table>	Assessment system Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)	Reference system (clinical diagnostic results)			Positive (+)	Negative (—)	Total	Positive (+)	198	4	202	Negative (—)	9	406	415	Total	207	410	617	Total compliance	95% confidence interval		97.89%	96.42%	98.87%
Assessment system Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)	Reference system (clinical diagnostic results)																									
	Positive (+)	Negative (—)	Total																							
Positive (+)	198	4	202																							
Negative (—)	9	406	415																							
Total	207	410	617																							
Total compliance	95% confidence interval																									
97.89%	96.42%	98.87%																								

(2) human throat swabs									
Assessment system Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)	Reference system (clinical diagnostic results)								
	Positive (+)	Negative(—)	Total						
Positive (+)	199	14	213						
Negative (—)	2	388	390						
Total	201	402	603						
<p>Sensitivity: 96.02%</p> <p>Spesitivity: 98.51%</p> <p>Accuracy: 97.68%</p> <p>Confidence interval analysis with a total compliance rate of 95%:</p> <table border="1" style="width: 100%;"> <tr> <td style="text-align: center;">Total compliance</td> <td colspan="2" style="text-align: center;">95% confidence interval</td> </tr> <tr> <td style="text-align: center;">97.68%</td> <td style="text-align: center;">96.14%</td> <td style="text-align: center;">98.72%</td> </tr> </table>				Total compliance	95% confidence interval		97.68%	96.14%	98.72%
Total compliance	95% confidence interval								
97.68%	96.14%	98.72%							
<p>2、 The consistency coefficient Kappa result of the diagnostic results between the assessment system and the reference system is below:</p> <p>Nasal swabs samples: Kappa (K) =0.9416;</p> <p>Throat swabs samples:Kappa (K) =0.9476;</p> <p>The assessment system can meet the current needs of clinical detection of the novel coronavirus 2019-nCoV antigen, and can be used to qualitatively detect the content of novel coronavirus 2019-nCoV antigen in human nasal swabs or throat swabs.</p>									
Verification unit:	The Key laboratory of Biological Emergency and Clinical POCT (Beijing)  Aug.17th,2020								

Note: The Key laboratory of Biological Emergency and Clinical POCT (Beijing) was jointly declared by Beijing Hotgen Biotech Co.,Ltd and institute of Microbiology of the Academy of Military Medical Sciences. It was announced on the website of the Beijing Municipal science & Technnology Commission on May 30, 2014.

Sensitivity verification of Novel Coronavirus 2019-nCoV

Antigen Test (Colloidal Gold)

Purpose

Use inactivated new coronavirus to evaluate the sensitivity of Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)

Experimental Materials

1. 1 batch of colloidal gold test paper;
2. Inactivated virus: 10^5 pfu/mL.

Experimental steps

Sample: Mixing ratio of sample diluent

Concentration number	Virus content in sample (pfu/mL)	Sample: Mixing ratio of sample diluent
1	0	1: 9
2	10^2	1: 9
3	2.5×10^2	1: 9
4	5×10^2	1: 9
5	10^3	1: 9
6	10^4	1: 9

1. After mixing the sample and diluent, incubate at room temperature for 1 min.
2. Take 100 μ L of sample and observe the result after 15min reaction.

Test results

Concentration number	Virus content in sample (pfu/mL)	Sample: Mixing ratio of sample diluent	Result
1	0	1: 9	-
2	10^2	1: 9	\pm
3	2.5×10^2	1: 9	+
4	5×10^2	1: 9	+
5	10^3	1: 9	++
6	10^4	1: 9	+++

In conclusion

Colloidal gold experiment results: 10^2 pfu/mL has a shallow band, negative without background, the sensitivity is 2.5×10^2 pfu/mL.

The Key Laboratory of Biological Emergency
and Clinical POCT (Beijing)
Aug. 17th, 2020



中国认可
检验
INSPECTION
CNAS IB0126

page 1 of 3 Pages

空运货物运输条件识别报告书

Certificate for Safe Transport of Air Cargo



证书编号: BN2009720700750002

物品名称: 新型冠状病毒(2019nCoV)抗原检测试剂盒(胶体金法)

Name of Goods: NOVEL CORONAVIRUS 2019-nCoV ANTIGEN TEST (COLLOIDAL GOLD)

签发日期: 2020-09-23

委托单位: 北京热景生物技术股份有限公司

Applicant:

北京信诺递捷运输咨询有限公司

SINO-Dangerous Goods Transportation Consultant Ltd.

电话: 010-64589142

网 址: www.chinasdg.cn

传真: 010-64580462

E-mail: public@chinasdg.cn

地址: 北京市顺义区北京空港物流基地物流园八街九号林吉大厦B505室

邮编: 101300

对外贸易经营者备案登记表

备案登记表编号: 01716790

统一社会信用代码: 91110115777090586H
 进出口企业代码: _____

经营者中文名称	北京热景生物技术股份有限公司		
经营者英文名称	Beijing Hotgen Biotech Co., Ltd.		
组织机构代码	_____	经营者类型 (由备案登记机关填写)	股份有限公司
住 所	北京市大兴区中关村科技园区大兴生物医药产业基地天富街9号9幢		
经营场所 (中文)	北京市大兴区中关村科技园区大兴生物医药产业基地天富街9号9幢		
经营场所 (英文)	9th Building, No.9 Tianfu St. Biomedical Base, Daxing District, Beijing, China		
联系电话	_____	联系传真	010-56528861
邮政编码	102600	电子邮箱	li.han@hotgen.com.cn
工商登记注册日期	2005-6-23	工商登记注册号	_____

依法办理工商登记的企业还须填写以下内容

企业法定代表人姓名	林长青	有效证件号	352202197609261014
注册资金	肆仟伍佰万元	(折美元)	

依法办理工商登记的外国(地区)企业或个体工商户(独资经营者)还须填写以下内容

企业法定代表人/个体工商户负责人姓名	_____	有效证件号	_____
企业资产/个人财产	_____	(折美元)	

备注	地址、变更, 原证号01224263 名称、经营者类型、注册资金变更, 原证号01224414
----	---

医疗器械生产许可证

许可证编号：京食药监械生产许20070010号

企业名称：北京热景生物技术股份有限公司

生产地址：北京市大兴区中关村科技园区大兴生物医药产业基地天富街9号9幢

法定代表人：林长青

生产范围：

2002版分类目录：II类：II-6840-3免疫分析系统，II-6840体外诊断试剂 III类：III-6840-3免疫分析系统，III-6840体外诊断试剂***

企业负责人：林长青

2017版分类目录：II类：II-22-04免疫分析设备 III类：III-22-15检验及其他辅助设备***

住 所：北京市大兴区中关村科技园区大兴生物医药产业基地天富街9号9幢

发证部门：



有效期限：至 2024 年 08 月 15 日

发证日期：2020 年 01 月 10 日

